


Root shaped conical implants for immediate multiple prosthesis

Uniko, one piece implant with its medium thread, allows the prosthetist to employ every connection of the Isomed Connector Bridge abutment System.



Uniko




Kits and instruments pag. 70
Prosthetic components pag. 91



Uniko

One piece implants with conical rotating connection bridge abutment



h = height
 ØA = maximum thread diameter
 ØB = apex diameter

Kits and instruments pag. 70
 Prosthetic components pag. 130

Connector Bridge Abutment Components	code	implant Ø	h	h shoulder	ØA	ØB
	UNIKO 3,5-10-SP1,5	3,5	10	1,5	3,5	2,65
	UNIKO 3,5-10-SP3	3,5	10	3,0	3,5	2,65
	UNIKO 3,5-12-SP1,5	3,5	12	1,5	3,5	2,65
	UNIKO 3,5-12-SP3	3,5	12	3,0	3,5	2,65
	UNIKO 3,5-14-SP1,5	3,5	14	1,5	3,5	2,65
	UNIKO 3,5-14-SP3	3,5	14	3,0	3,5	2,65
Connector Bridge Abutment Components	code	implant Ø	h	h shoulder	ØA	ØB
	UNIKO 4-10-SP1,5	4,0	10	1,5	4,0	3,1
	UNIKO 4-10-SP3	4,0	10	3,0	4,0	3,1
	UNIKO 4-11,5-SP1,5	4,0	11,5	1,5	4,0	3,1
	UNIKO 4-11,5-SP3	4,0	11,5	3,0	4,0	3,1
	UNIKO 4-13-SP1,5	4,0	13	1,5	4,0	3,1
	UNIKO 4-13-SP3	4,0	13	3,0	4,0	3,1
Connector Bridge Abutment Components	code	implant Ø	h	h shoulder	ØA	ØB
	UNIKO 4,5-10-SP1,5	4,5	10	1,5	4,5	3,6
	UNIKO 4,5-10-SP3	4,5	10	3,0	4,5	3,6
	UNIKO 4,5-11,5-SP1,5	4,5	11,5	1,5	4,5	3,6
	UNIKO 4,5-11,5-SP3	4,5	11,5	3,0	4,5	3,6
	UNIKO 4,5-13-SP1,5	4,5	13	1,5	4,5	3,6
	UNIKO 4,5-13-SP3	4,5	13	3,0	4,5	3,6
Connector Bridge Abutment Components	code	implant Ø	h	h shoulder	ØA	ØB
	UNIKO 5-10-SP1,5	5,0	10	1,5	5,0	4,1
	UNIKO 5-10-SP3	5,0	10	3,0	5,0	4,1
	UNIKO 5-11,5-SP1,5	5,0	11,5	1,5	5,0	4,1
	UNIKO 5-11,5-SP3	5,0	11,5	3,0	5,0	4,1
	UNIKO 5-13-SP1,5	5,0	13	1,5	5,0	4,1
	UNIKO 5-13-SP3	5,0	13	3,0	5,0	4,1

Surgical protocol

	CMU-450	SF00	SF 20-stop h 10-11,5-13-14,5	SF 35 h 10-12-14	SF 40 h 10-11,5-13-14,5	SF 45 h 10-11,5-13-14,5	SF 50 h 10-11,5-13
	Handpiece punch	Pilot drill	Conical drills				
Ø 3,5	•	•		•			
Ø 4,0	•	•		•	•		
Ø 4,5	•	•		•	•	•	
Ø 5,0	•	•		•	•	•	•
For D1/D2 bone							
Ø 3,5	•	•	•*	•			
Ø 4,0	•	•		•*	•		
Ø 4,5	•	•		•	•*	•	
Ø 5,0	•	•		•	•	•*	•
For D3/D4 bone							
Ø 3,5	•	•		•			
Ø 4,0	•	•		•	•		
Ø 4,5	•	•		•	•	•	
Ø 5,0	•	•		•	•	•	•

*Fixtures can be placed with an intermediate drill or bone expanders in D4 bone, achieving an optimal primary stability. Trans-mucous shoulder of 1,5 or 3 mm height can be placed on the gum.

Protocols and sequences are just suggested with an illustrative purpose. It's up to the surgeon to select the best surgical option for the anatomy of the patient.